

# Skin Temperature Prediction in Lower Limb Prostheses

Neha Mathur\*, *Student Member, IEEE*, Ivan Glesk, *Senior Member, IEEE* and Arjan Buis

**Abstract**— Increased temperature and perspiration within a prosthetic socket is a common complaint of many amputees. The heat dissipation in prosthetic sockets is greatly influenced by the thermal conductive properties of the socket and interface liner materials. These materials influence the body's temperature regulation mechanism and might be the reason for thermal discomfort in prosthetic sockets. Monitoring interface temperature at skin level is notoriously complicated. The problem might be considered notorious because embedding wires and sensors in an elastomer eventually results in elastomer failures because of the high strain induced when donning a liner (amputees roll the liners onto their limbs). Another reason is because placing sensors and wires directly against the skin could cause irritation and chaffing over just a short period of time. We describe a route wherein if, the thermal properties of the socket & liner materials are known, the in-socket residual limb temperature could be accurately predicted - by monitoring the temperature between socket and liner rather than skin and liner using the Gaussian Processes technique.

**Index Terms**—Gaussian Process for Machine Learning, Lower Limb Prosthetics, Modeling, Temperature.

## I. INTRODUCTION

MONITORING and predicting the residual limb skin health in lower limb amputees is of principal importance as the socket of the prosthesis creates a warm and humid micro environment that encourages growth of bacteria and skin breakdown. Elevated residual limb skin temperature is considered one of the major factors that could affect the health of soft tissues in that region. Studies on temperature within the prosthetic sockets of trans-tibial prosthetic users have been described by Peery et al [13], [14]. They investigated the in-socket temperature of five trans-tibial amputees at fourteen different locations on the residual limb and at four different stages i.e. donning; steady state resting; initial walking and

This work was supported by the Engineering and Physical Sciences Research Council under the Doctoral Training Grant and the Centre for Excellence in Rehabilitation Research (CERR). Also, support for the climate chamber was given by University of Glasgow. Asterisk indicates corresponding author.

\*Neha Mathur is with the Electronic and Electrical Engineering Department, University of Strathclyde, Glasgow G11XW, UK (email: neha.mathur@strath.ac.uk).

Ivan Glesk is with the Electronic and Electrical Engineering Department, University of Strathclyde, Glasgow G11XW, UK (email: ivan.glesk@strath.ac.uk).

Arjan Buis is with the Department of Biomedical Engineering, University of Strathclyde, Glasgow G40NW, UK (email: arjan.buis@strath.ac.uk).

steady state walking. Their results indicated that the thermal dissipation characteristics of socket and liner restrict heat loss from the residual limb and the temperature increase is larger in areas where there is more muscle bulk. It was also seen that different socket and liner materials affect the temperature increase in the residual limb differently. Also these temperature rises were different between patients. The impact of environmental factors has also been investigated by Klute et al [26] by looking at the in-socket residual limb temperature at four locations throughout a whole day. Simultaneously, they recorded environmental temperature, humidity and also the activity of the patient. It was found that in-socket residual limb temperature increased gradually throughout the day and an increase in activity caused a further increase in temperature. Also, the environmental humidity and temperature influence the perception of whole body and residual limb thermal comfort. Limitations to this study were small sample size and the variation of prosthetic liners and socket materials between patients. The population on which these studies were done was small. Hence, a generalized statement about the temperature of residual limb skin cannot be made.

Limited research has been done on the thermal behavior of the socket and liner materials and their effect on stump skin temperatures. Klute et al [8] investigated the thermal conductivity of different liner and socket materials. Their investigations disclosed that there was a large variation in the thermal conductivity of liner materials. Whereas the prosthetic socket materials – thermoplastic and carbon fiber had similar thermal conductivities. They concluded that liner material selection has a considerable effect on the residual limb skin temperature as compared to the thermoplastic and carbon fiber socket materials which has little effect. The results presented in [8] suggested that some prosthetic components can act as a barrier to conductive heat transfer due to low thermal conductivity. This implies that different liner and socket materials produce a different thermal environment and hence can lead to different residual limb skin temperatures.

This leads to a hypothesis that if the thermal properties of the socket & liner materials are known then the in-situ skin temperature could be predicted by monitoring between socket and liner. The purpose of this is to assess whether or not a temperature measurement device can accurately measure the temperature of the residual limb when it is placed either on the inner or outer surface of a prosthetic socket. If that is achievable then the monitoring of the residual limb skin temperature can be done without undesirable contact of any

temperature sensor with the skin thus avoiding any increased skin irritation.

Our approach was to determine an accurate mathematical model using the Gaussian processes for machine learning (GPML) to predict the residual limb skin temperature of the amputee. The temperature measurement device when placed directly in contact with the skin would give the most accurate reading. However, this could create practicality issues with normal prosthetic use in a domestic situation such as, protruding lead wiring, consistent positioning of sensors and possible skin irritation. Embedding sensors and wires in to the hard prosthetic socket is possible during the manufacturing process for prosthetic sockets. This would eliminate the practicality issues as described earlier. In addition, no damage to the device would occur through donning and doffing and the longevity of the device would not be impaired.

The aims of this project are to investigate how differing activity levels and ambient temperatures influence the in-socket temperature. It would be then investigated that whether the predictive model developed is accurate enough and its response time and repeatability are also sufficient.

## II. METHOD

To investigate the correlation between the position of thermocouples (skin and in-socket), one trans-tibial traumatic amputee was recruited to perform in a 35 minute laboratory protocol. The investigation was implemented following ethical approval granted by the University of Strathclyde Ethics Committee (Ref UEC13/04).

The subject was a 68 year-old male who weighed 70 kg and wore a 6mm Otto Bock HealthCare GmbH, Technogel liner with a resin laminate socket. For the experiment the subject was dressed in shorts and t-shirt. Hence, it should be noted that there was no extra clothing layer on the prosthesis. To monitor and record the skin and in-socket temperatures, four K-type thermocouples via a data logger (type HH1384; Omega Engineering) were used. Two thermocouples were taped onto the residual limb in lateral and medial position. The other two thermocouples were put on the corresponding positions on the liner (in-socket). This is indicated in Fig 1. Data from the four channels was simultaneously collected at 0.5Hz at a predetermined ambient temperature.

After the thermocouple heads were secured with tape, the prosthesis was donned with the thermocouple wires exiting the proximal edge of the socket. The subject was asked to complete the following protocol: resting (sitting) for 10 minutes, walk at self-selected pace of 0.62 metres/second on a treadmill for 10 minutes, and finally rest for 15 minutes. The residual limb skin and the socket temperatures were sampled at 0.5Hz for the entire 35 minutes protocol. For analysis purposes, three steady state periods were defined as the last minute of each period: initial rest, walking, and final rest. The temperature profile of the residual limb skin and liner were analysed at different ambient temperatures to see how closely they are correlated. This study was conducted in Scotland for the Spring/Summer profile where the ambient temperature ranges from approximately 10°C-25°C. Hence, the

temperatures from this range were picked. We conducted the experiment at 10°C and then repeated for 15°C, 20°C, and 25°C (Data Set A). The experiments were conducted again after a time span of two months to confirm the influence of ambient temperature on the residual limb skin temperature (Data set B). All experiments were conducted in a climate controlled chamber with zero wind velocity and 40% humidity level.



Fig. 1. Placement of the thermocouples on the lateral and medial side of residual limb skin and its corresponding positions on the liner of the trans-tibial amputee


## III. RAW DATA

The temperature profiles of the liner and residual limb skin at ambient temperatures of 10°C, 15°C, 20°C, and 25°C from set B are shown in Fig 2, Fig 3, Fig 4, and Fig 5.


From the studies on the amputee subject, it is seen that donning causes a moderate temperature increase, walking causes a significant increase, and the rest periods following activities must be substantially long to return the limb to temperatures before donning the prosthesis. Fig 2 and Fig 3 indicate that the residual limb temperature profile for ambient temperatures of 10°C and 15°C has a similar pattern of being steady throughout the experiment. However, this temperature profile of the residual limb is significantly different from that at ambient temperatures of 20°C and 25°C as indicated in Fig 4 and Fig 5.

Both the lateral and medial residual limb skin temperatures showed a steady increase in the temperature throughout the experiment. After the end of the experiment the temperatures in both lateral and medial side were 2.1°C higher than the starting.


It can be seen from these studies on amputee subject that the residual limb skin temperature behavior is a function of ambient temperature. This reflects that though the human body self regulates to maintain a stable internal environment despite changes in the external environment, in case of prosthetic users there are layers of liner and socket materials which inhibit the body's ability to thermoregulate effectively.




(a)



(a)




(b)




(b)

Fig. 2. Profiles of residual limb skin and liner temperature at ambient temperature 10°C on (a) lateral side (b) medial side


Fig. 4. Profiles of residual limb skin and liner temperature at ambient temperature 20°C on (a) lateral side (b) medial side




(a)



(a)



(b)



(b)

Fig. 3. Profiles of residual limb skin and liner temperature at ambient temperature 15°C on (a) lateral side (b) medial side

Fig. 5. Profiles of residual limb skin and liner temperature at ambient temperature 25°C on (a) lateral side (b) medial side

Since, the temperature profiles of the residual limb are almost similar for the ambient temperature pairs of 10°C, 15°C and 20°C, 25°C, the Gaussian predictive model at ambient temperatures of 10°C and 25°C are only discussed in this study.

#### IV. GAUSSIAN PROCESS FOR MACHINE LEARNING

The data from the data logger indicated that at any given ambient temperature, the trace of the liner temperature follows that of the residual limb skin as in Fig 2 and Fig 3. This suggested a possibility to model the liner temperature as a function of skin temperature and create a mathematical model of the same. Processing was performed with custom developed software (using Matlab®, Mathworks). The model designed takes the liner temperature as the input  $x$  and the predicted output is the residual limb skin temperature  $y$ .

Predictive model was developed using the Gaussian Process for Machine Learning [30]. Similar with other regression methods, the goal of Gaussian Process regression is to infer a continuous function  $f(x)$  from a training set of input-output pairs in a supervised learning context. The key assumption in Gaussian Process modelling is that our data can be represented as a sample from a multivariate Gaussian distribution. A Gaussian process model infers a joint probability distribution over all possible outputs for all inputs. This form enables the implementation of Bayesian framework in a simple way [28], [29]. Bayes' theorem states that the posterior probability of a condition is given by the product of the prior probability and the likelihood in the light of the evidence. This can be written as

$$P(B|A, H) = \frac{\text{likelihood} \cdot \text{prior}}{\text{evidence}} = \frac{P(A|B, H) \cdot P(B|H)}{P(A|H)} \quad (1)$$

$P(B|A, H)$  is the posterior probability that statement  $B$  is true, given that condition  $A$  is observed and that hypothesis  $H$  is correct.  $P(A|B, H)$  is the probability of observing  $A$  if  $B$  is true and  $H$  is correct, which is called the likelihood.  $P(B|H)$  is the prior probability of  $B$  being true, without having made any observations.  $P(A|H)$  is the evidence: the probability of observing  $A$  if hypothesis  $H$  is correct [29].

The inference of a joint probability distribution function involves deducing a number of quantities called the hyperparameters  $\Theta$ . These hyperparameters are an indication of the precision and relevance of the input parameters for predicting the output. Thus, the aim in a Gaussian process model is to choose model parameters for which the probability of the training data is maximized [28].

If the training data contains  $N$  points comprising of their set of targets  $y_N$  with their corresponding input  $x_N$ , then the Gaussian model is defined by  $N$  dimensional covariance matrix  $C_N$ . The covariance matrix is basically indicative of the closeness to each other outputs for different inputs, taking into account the model parameters. This allows predictions of

outputs  $y_*$  to be made, based on the difference between the new inputs  $x_*$  and those seen in the training data. Each element of  $C_N$  is defined by covariance function  $C_f$ , which is a function of inputs and hyperparameters [28], [29], [31]. For the element  $ij$  in covariance matrix  $C_{ij} = C_f(x_i, x_j, \Theta)$ . The covariance function can be user defined. In this study, the covariance function that was used was

$$C_f = \theta_1 e^{-\frac{(x_i - x_j)^2}{2l^2}} + \sigma_n^2 \delta_{ij} \quad (2)$$

where the set of hyperparameters  $\Theta = \{\theta_1, l, \sigma_n\}$  and  $\delta_{ij}$  is a delta function whose value is zero for all  $i \neq j$ . The first term in the above equation allows the closeness of two outputs to be related to the closeness of the inputs. The length scale  $l$  for an input parameter indicates how much the output will vary relative to changes in an input. The initial values of the hyperparameters were selected as [0.1, 1, 1]. To prepare for Gaussian process regression, we calculate the covariance function, (2), for all possible combinations of these points, summarizing in the matrices in (3), (4), and (5):

$$C_N = \begin{bmatrix} C_f(x_1, x_1) & C_f(x_1, x_2) & \dots & C_f(x_1, x_N) \\ C_f(x_2, x_1) & C_f(x_2, x_2) & \dots & C_f(x_2, x_N) \\ \vdots & \vdots & \ddots & \vdots \\ C_f(x_N, x_1) & C_f(x_N, x_2) & \dots & C_f(x_N, x_N) \end{bmatrix} \quad (3)$$

$$C_{N*} = [C_f(x_*, x_1) \quad C_f(x_*, x_2) \quad \dots \quad C_f(x_*, x_N)] \quad (4)$$

$$C_{N**} = C_f(x_*, x_*) \quad (5)$$

Since the key assumption in Gaussian Process modeling is that the data can be represented as a sample from a multivariate Gaussian distribution, we have

$$\begin{bmatrix} y \\ y_* \end{bmatrix} \sim \left( 0, \begin{bmatrix} C_N & C_{N*}^T \\ C_{N*} & C_{N**} \end{bmatrix} \right) \quad (6)$$

where T indicates the matrix transposition. The conditional probability  $p(y_*|y)$ : "given the data, how likely is certain prediction for  $y_*$ ", follows a Gaussian distribution as in (7)

$$y_*|y \sim (C_{N*} C_N^{-1} y, C_{N*} C_N^{-1} C_{N*}^T - C_{N**}) \quad (7)$$

The reliability of the regression depends upon on covariance function and in turn the hyperparameters. Typically, we would not a priori know the values of the hyperparameters. To get the optimal hyperparameters, (1) becomes

$$P(\Theta | y_N, x_N, C_f) = \frac{P(y_N | x_N, C_f, \Theta) \cdot P(\Theta)}{P(y_N | x_N, C_f)} \quad (8)$$

Referring to (8), it is apparent that the evidence is independent of hyperparameters and is constant for a given dataset. To find the optimal hyperparameters, the posterior probability is maximized as the prior maybe non-informative. This corresponds to minimizing the log marginal likelihood (nlml) as in (9). Hence, for a particular training set and covariance function, the Gaussian process would select the

best hyperparameters that give the best predictions for training data [29], [31].

$$\log p(y|x, \Theta) = -\frac{1}{2}y^T C_N^{-1}y - \frac{1}{2}\log|C_N| - \frac{n}{2}\log 2\pi \quad (9)$$

Several multivariate optimization algorithms can be utilized to calculate the hyperparameters, such as Laplace's approximation, Markov Chain Monte Carlo sampling, Kullback-Leibler optimal approximation or the Variational Bayes approximation. In the present study, kernel hyperparameters were optimized by the exact inference technique. For real-valued outputs, it combines the Gaussian process prior with a Gaussian likelihood and perform an exact posterior inference in closed form.

### V. MODEL GENERATION AND PREDICTION

The aim of the model is to predict the residual limb skin temperature from the liner temperature. Hence, the input to the model is the liner temperature and the output of the model would be the residual limb skin temperature. To test the predictive capability of a model, it is trained on one set of data and tested on previously unseen data. It is seen from Section II that the skin temperature is dependent on ambient temperature. Hence, individual Gaussian process models for the lateral and medial side of the residual limb were designed, using the principle as described in the previous section for ambient temperatures of 10°C and 25°C. Consider the ambient temperature 10°C, first the lateral side model was trained on different scenarios to investigate the optimal training required. Table 1 presents the different testing and training cases presented to the model. Initially the model was trained was by 250 data points from set A. The training points are inclusive of the three stages of protocol (initial rest, walking, final rest) and are picked in the ratio of time intervals used for the respective protocol periods. Hence, the first 70 points from initial rest, first 70 points from walking and first 110 points from the final rest were taken for training. The predictive capability of the model is gauged by computing the training error, test error and the normalized log likelihood by testing with 100 data points (not seen by the model during training) from set A which are again drawn in the ratio of time intervals used for the respective protocol periods. This process is continued by increasing the number of training data points from set A. The normalized log likelihood for each set of test data is also given, calculated by dividing the value of marginal likelihood by the number of points in the test set. It can be seen that as the number of training points increase, the error value decreases and the likelihood of the data increases. This implies that with greater training points, the new model either predicts data closely or has higher confidence due to a higher density of training points.

When the model is trained on all the values of set A and tested with 100 points (randomly picked from set A), the training error does decrease while the likelihood also decreases slightly. This might be an indication that too much of training to the model might lead to over fitting the data. Next the model trained on the entire set A was tested on 100 points in set B which were unseen by the model. The results indicated that the test error has a substantial increase but the

likelihood function is still consistent with the uncertainty predictions of the Gaussian model. This may be because the data set B is similar to data set A, leading to points closer to the smooth relationship predicted by the Gaussian process. It was then deduced that when the model is trained on whole data set A and hyperparameters optimized in (9) then its predictions for set B lie in the 95% confidence interval ( $\pm 2$  standard deviations). This is indicated in Fig 6 and Fig 7.

TABLE I  
SUMMARY OF GPML FOR VARIOUS TESTING AND TRAINING SCENARIOS

Training Set	Test Set	Training error (rms)	Test error (rms)	Number of training points	Number of test points	Normalized log likelihood
A	A	0.142	0.142	250	100	0.786
A	A	0.0924	0.0924	500	100	0.893
A	A	0.0913	0.0913	750	100	0.897
A	A	0.0910	0.0910	1050	100	0.895
A	B	0.0910	0.102	1050	100	0.896

After hyperparameter optimization, the covariance hyperparameters for the lateral side at 10°C were [-1.27, 2.92, -1.41] and the likelihood hyperparameter was -1.79. The final negative log marginal likelihood (optimized) was 285.84. Table 2 presents the hyperparameters for the predictive model at different ambient temperatures.

TABLE II  
SUMMARY OF HYPERPARAMETERS FOR PREDICTIVE GAUSSIAN MODEL

Scenario	Optimized hyperparameters	Initial nml	Final nml	Initial likelihood	Final likelihood
Lateral side at 10°C	[-1.27, 2.92, -1.41]	786.11	285.84	-1.78	-1.79
Medial side at 10°C	[-1.20, -0.86, 1.29]	747.63	352.61	-2.17	-2.21
Lateral side at 25°C	[-1.86, -1.37, -1.04]	538.43	323.78	-1.78	-1.80
Medial side at 25°C	[-1.35, -1.57, 1.06]	520.86	109.68	-1.96	-1.99

Similar process was done for the medial side of residual limb skin. After the predictive algorithm was formulated, the actual and predicted data is then averaged after every 5 samples to create a single mean, and done till the end of all data points. This helps in smoothening out short-term fluctuations and highlighting the longer-term trends.




Fig. 6. Illustration of prediction with Gaussian Process regression for ambient temperature of 10°C at (a) lateral side (b) medial side. The test data points are given by crosses. The shaded area represents the point wise 95% confidence region of the predictive distribution.

## VI. DISCUSSION

From the raw data it can be noted that for the ambient temperature of 10°C, the lateral skin temperature was almost same at the start and end of the experiment. However, the medial residual limb skin temperature followed a steady profile but the temperature at the end of the experiment was 1°C lower than the start. At ambient temperature of 25°C, both the lateral and medial residual limb skin temperatures showed a steady increase in the temperature throughout the experiment. After the end of the experiment the temperatures in both lateral and medial side were 2.1°C higher than the starting. It can be seen from Fig 2 and Fig 3 that since the volume of the data set are big, the overall trend of the temperature profile is difficult to gauge. Time averaging of 5 seconds done on the recorded data helps in identifying the trend better and improves the joint probability function with an enhanced fit of the Gaussian to the data; so that more points are more accurately predicted. The actual skin temperature obtained by the Gaussian predictive model is shown in Fig 8 and Fig 9 for two very different ambient temperatures of 10°C and 25°C, respectively.




Fig. 7. Illustration of prediction with Gaussian Process regression for ambient temperature of 25°C at (a) lateral side (b) medial side. The test data points are given by crosses. The shaded area represents the point wise 95% confidence region of the predictive distribution.

For both these experiments done at two different ambient temperatures, our predictive model using Gaussian process provides a simple, effective, practical and probabilistic approach to determine the unknown skin temperature of the subject within the prosthetic device from the actual liner measurements. The predictive model we developed leads to results which are in 95% confidence interval which translates to an accuracy of  $\pm 0.8^\circ\text{C}$ . However this approach has certain limitations as well. Although this study was conducted on one amputee subject over a number of times to verify the influence of ambient temperatures on the in-socket temperatures, there is a need to extend it on a greater population in order to define a generic behaviour. Since the residual limb temperature profile varies with changes in environmental temperatures, the Gaussian model has to be trained with individual datasets corresponding to changes in the ambient temperatures.

## VII. CONCLUSION

The challenge of non-invasively monitoring the residual limb skin temperature has been addressed in this research work. This study which was conducted on a subject performing various tasks in an environmental chamber at different ambient temperatures clearly indicated that the residual limb skin temperature and the liner temperature are

majorly affected by both the ambient temperature and the activity level of the subject. The obtained data from the experimentation at ambient temperatures of 10 °C and 25°C was then used to develop a predictive model using Gaussian processes for machine learning. GPML is a non-parametric approach, which harnesses the training data provided to determine the underlying function. The Gaussian model was individually trained for each of the ambient temperatures on which the tests were done. With 5 seconds time averaging in order to suppress random fluctuations the developed model provides results which lie in the 95% confidence interval when predicting subject skin temperature within the prosthesis from the liner temperature measurements. Future scope of the work includes comparing the prediction accuracy of the Gaussian Process model with other mathematical models like the Adaptive Network-Based Fuzzy Inference System (ANFIS) using Support Vector Regression (SVR).




Fig. 8. The predicted residual limb skin temperature from the time averaged Gaussian Process Model is shown along with the actual skin temperature at lateral and medial sides in (a) and (b) respectively at ambient temperature of 10°C





Fig. 9. The predicted residual limb skin temperature from the time averaged Gaussian Process Model is shown along with the actual skin temperature at lateral and medial sides in (a) and (b) respectively at ambient temperature of 25°C

### REFERENCES

- [1] K.Hagberg and R.Branemark, "Consequences of non-vascular transfemoral amputation: a survey of quality of life, prosthetic use and problems," *Prosthetics and Orthotics International*, vol. 25, no. 3, pp. 186-194, Dec. 2001.
- [2] F.T. Hoaglund, H.E. Jergesen, L. Wilson, L.W. Lamoreux, and R. Roberts, "Evaluation of problems and needs of veteran lower-limb amputees in the San Francisco Bay Area during the period 1977-1980," *Journal of rehabilitation R&D / Veterans Administration, Department of Medicine and Surgery, Rehabilitation R&D Service*, vol. 20, no.1, pp.57-71, Jul. 1983.
- [3] S. Derler and L. Gerhardt, "Tribology of Skin: Review and Analysis of Experimental Results for the Friction Coefficient of Human Skin," *Tribol Lett.*, vol.45, no.1, pp.1-27, Jan. 2012.
- [4] J.E. Sanders, B.S.Goldstein, and D.F. Leotta, "Skin response to mechanical stress adaptation rather than breakdown – a review of the literature," *Journal of Rehabilitation Research and Development*, vol 32, no.3, pp.214-226, Oct. 1995.
- [5] A.F.T. Mak, M. Zhang, and D.A. Boone, "State-of-the-art research in lower-limb prosthetic biomechanics-socket interface: A review," *Journal of Rehabilitation Research and Development*, vol 38, no.2, pp.161-173, Mar./Apr. 2001.
- [6] S. Levy, "Skin problems of the leg amputee," *Journal of Prosthetics & Orthotics*, vol.4, no.1, pp.37-44, Apr. 1980.
- [7] Barnes R. Thermography. *Annals of The New York Academy of Sciences*. 1964;121:34-48.
- [8] G.K. Klute, G.I. Rowe, A.V. Mamishev and, W.R. Ledoux, "The thermal conductivity of prosthetic sockets and liners," *Prosthetics and Orthotics International*, vol.31, no.3, pp.292-299, Sep.2007.
- [9] R. Emrich and K. Slater, "Comparative analysis of below-knee prosthetic socket liner materials," *Journal of Medical Engineering & Technology*, vol.22, no.2, pp.94-98, Jan.1998.
- [10] M. Gerschutz, M. Haynes, D. Nixon, and J. Colvin, "Strength evaluation of prosthetic check sockets, copolymer sockets, and definitive laminated sockets," *Journal of Rehabilitation and Research Development*, vol.49, no.3, pp.405-426, May 2012.
- [11] M. Gerschutz, M. Haynes, D. Nixon, and J. Colvin, "Tensile strength and impact resistance properties of materials used in prosthetic check sockets, copolymer sockets, and definitive laminated sockets," *Journal of Rehabilitation Research & Development*, vol.48, no.8, pp.987-1004, Nov. 2011.
- [12] D. Taylor, "Lower Limb Prosthetics - Mechanical Testing Prosthetic Sockets," in *Academy of Orthotist and Prosthetists Annual Meeting and Scientific Symposium*, Orlando, Florida, Mar.16-19, 2005.
- [13] E.A. Huff, W.R. Ledoux, J.S. Berge, and J.K. Klute, "Measuring Residual Limb Skin Temperatures at the Skin-Prosthesis Interface," *Journal of Prosthetics and Orthotics*, vol.20, no.4, pp.170-173, Oct. 2008.
- [14] J.T. Peery, W.R. Ledoux, and G.K. Klute, "Residual-limb skin temperature in transtibial sockets," *Journal of Rehabilitation Research and Development*, vol.42, no.2, pp.147-154, Mar./Apr. 2005.
- [15] A. Shitzer and R. Eberhart, *Heat Transfer in Medicine and Biology: Analysis and Applications*. USA: Plenum Press, 1983.

- [16] Y. Houdas, and E. Ring, *Human Body Temperature Its Measurement and Regulation*. USA: Plenum Press, 1982.
- [17] R. Frykberg, T. Zgonis, D. Armstrong, V. Driver, G. Giurini, S. Kravitz, et al., "Diabetic Foot Disorders: A Clinical Practice Guideline (2006 Revision)," *The Journal of Foot and Ankle Surgery*, vol.45, no.5, pp.1-66, Sep./Oct. 2006.
- [18] L.A. Sokolnicki, N.A. Strom, S.K. Roberts, S.A. Kingsley-Berg, A. Basu, and A. Charkoudian, "Skin blood flow and nitric oxide during body heating in type 2 diabetes mellitus," *Journal of Applied Physiology*, vol.106, no.2, pp.566-570, Feb. 2009.
- [19] N. Charkoudian, "Skin blood flow in adult thermoregulation: how it works, when it does not and why," *Mayo Clinic Proceedings*, vol.78, no.5, pp.603-612, May 2003.
- [20] J. Petrofsky, "The Effect of Type-2-Diabetes-Related Vascular Endothelial Dysfunction on Skin Physiology and Activities of Daily Living," *Journal of Diabetes Science and Technology*, vol.5, no.3, pp.657-667, May 2011.
- [21] A.E. Caballero, S. Arora, R. Saouaf, S.C. Lim, P. Smakowski, J.Y. Park, G.L. King, F.W. LoGerfo, E.S. Horton, and A.Veves, "Microvascular and macrovascular reactivity is reduced in subjects at risk for type 2 diabetes," *Diabetes*, vol.48, no.9, pp.1856-1862, Sep. 1999.
- [22] J. Petrofsky, S. Lee, C. Patterson, M. Cole, and B. Stewart, "Sweat production during global heating and during isometric exercise in people with diabetes," *Medical Science Monitor*, vol.11, no.11, pp.CR515-CR521, Nov. 2005.
- [23] D. Wick, S. Roberts, A. Basu, P. Sandroni, R. Fealey, D. Sletten, and N. Charkoudian, "Delayed threshold for active cutaneous vasodilation in patients with type 2 diabetes mellitus," *Journal of Applied Physiology*, vol.100, no.2, pp.637-641, Feb. 2006.
- [24] K.S. Maluf, R. Morley, Jr., E. Richter, J. Klaesner, and M. Mueller, "Monitoring In-Shoe Plantar Pressures, Temperature, and Humidity: Reliability and Validity of Measures From a Portable Device," *Archives Physical and Medical Rehabilitation*, vol.82, no.8, pp.1119-1129, Aug. 2001.
- [25] G. Klute, J. Berge, E. Huff, and W. Ledoux, "The Effect of Rest and Exercise on Residual Limb Temperature," in *American Academy of Orthotists and Prosthetists Annual Meeting and Scientific Symposium*, Chicago, Illinois, Mar.1-9, 2006.
- [26] G. Klute, E. Huff, and W. Ledoux, "In-Socket Skin Temperature and Perception of Comfort Over A Whole Day," in *American Academy of Orthotists and Prosthetists Annual Meeting and Scientific Symposium*, Seattle, Washington, 2007.
- [27] J. Peery, G. Klute, J. Blevins, and W. Ledoux, "A Three-Dimensional Finite Element Model of the Transtibial Residual Limb and Prosthetic Socket to Predict Skin Temperatures," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol.14, no.3, pp.336-343, Sep. 2006.
- [28] C. K. I. Williams and C. E. Rasmussen, "Gaussian Processes for regression," *Neural Information Processing Systems*, vol 8, pp 514-520, MIT Press, Cambridge, MA, 1996.
- [29] C. E. Rasmussen and C. K. I. Williams, *Gaussian processes for machine learning. Adaptive computation and machine learning*. Cambridge: MIT Press, 2005.
- [30] C. E. Rasmussen and C. K. I. Williams. (2013, January). Gaussian Processes for Machine Learning software. [Online]. Available GAUSSIANPROCESS.ORG/GPML/
- [31] M.N. Gibbs, "Bayesian Gaussian processes for regression and classification," Ph.D. dissertation, University of Cambridge, 1998. Obtainable from <http://wol.ra.phy.cam.ac.uk/mng10/GP/GP.html>



**Neha Mathur** received the B.E. degree in electrical engineering from Bharati Vidyapeeth University, India, in 2005 and the M.S. degree from the University of New Hampshire, USA, in 2008. She is currently pursuing the Ph.D. degree in electronic and electrical engineering at University of Strathclyde, Glasgow.

From 2008 to 2011, she joined Corus Steel, Port Talbot, UK where she was a Process Control and Automation Engineer. Her research interest includes

development of sensor systems for data monitoring, analysis and prediction of residual limb health in prosthetics.



**Ivan Glesk** (M'01–SM'01) received the Ph.D. degree in quantum electronics and optics from Comenius University, Slovak Republic, in 1989 and the D.Sc. degree from the Slovak Academy of Sciences. In 1986, he joined the Department of Experimental Physics, Comenius University, where he later became a Professor of physics, and was engaged in

the research in areas of nonlinear optics, laser physics, and using LIDAR methods for remote sensing the atmosphere. In 1991, he joined Princeton University, USA where he was Manager of the Lightwave Communication Research Laboratory and Senior Research Scholar in the Department of Electrical Engineering. In 2007, he joined University of Strathclyde, UK, as a Professor of Broadband Communication Systems in the Department of Electronic and Electrical Engineering. His current research interests include biophotonics, sensors, applications of ultra-short pulses in data communication, ultrafast optical signal processing, and optical interconnects. He is author and co-author of more than 290 publications, 16 book chapters and holds five patents. He is on the Editorial Board of the *International Journal of Optics* and *Optica Applicata* and is a recipient of the International Research and Exchanges Board (IREX) Fellowship.



**Arjan Buis** is a Senior Research Fellow at the Department of Biomedical Engineering in University of Strathclyde, Glasgow. He started his PhD education at the University of Strathclyde, Glasgow and graduated in 1997. His current research relates to improving our understanding of the biomechanical mechanisms that contribute to the generation and control of load transfer forces dealing with the subject "where man meets machine" and especially the area of prosthetic socket fit. Examples of this are evident in his research related to the capture of proof of the dynamic mechanical environment at the stump-socket interface, within the deeper structures of the stump and the overall impact of this approach on amputee gait performance and acceptance. Beside the biomechanical interests he is also developing a portfolio in relation to component design related topics and includes; "Smart" materials applicable in rehabilitation technology and development of specific prosthetic components using alternative materials and composites. In addition he is leading a research group dealing with integrated intelligent sensor systems to enable built-in data monitoring of prosthetics and orthotics as a means to maintain optimum performance and quality of life.